



Technical Specifications

Affinity<sup>2.0</sup> / Equinox<sup>2.0</sup>

**Hearing Aid Analyzer**

---



## Included and Optional Parts

AC440	REM440	HIT440
<p><b>Included parts:</b></p> <ul style="list-style-type: none"> <li>• Affinity2.0 AC440 CD</li> <li>• OtoAccess™ database CD</li> <li>• TDH39 Audiometric headset or DD45 Audiometric headset</li> <li>• MTH400 Headset</li> <li>• EMS400 Talk back microphone</li> <li>• B81 Bone conductor</li> <li>• APS3 Patient response button</li> <li>• Standard USB cable</li> <li>• Power cable 120 or 230V</li> <li>• Mouse pad</li> <li>• Instructions for Use document</li> </ul> <p><b>Optional parts:</b></p> <ul style="list-style-type: none"> <li>• DAK70 Audiometer keyboard with live voice mic.</li> <li>• Earphone 3A insert earphones (5As may be substituted)</li> <li>• IP30 insert earphones</li> <li>• B81 Bone Conductor</li> <li>• B71 Bone Conductor</li> <li>• ACC60 Affinity2.0 carrying case</li> <li>• CIR22 Insert masking earphones</li> <li>• Audiocup enclosures</li> <li>• Peltor noise excluding headset</li> <li>• HDA280 Audiometric headset</li> <li>• HDA300 Audiometric headset</li> <li>• KOSS R80 high frequency headset</li> <li>• AP70 Power amplifier 2x70 Watt</li> <li>• SP90 Loudspeaker</li> <li>• SP85A Loudspeaker</li> <li>• SP90A Loudspeaker</li> <li>• AFC8 Sound cabin installation panel</li> <li>• Optical USB 1.1 isolation extension cable</li> </ul> <p><b>Optional special tests:</b></p> <ul style="list-style-type: none"> <li>• High Frequency audiometry (HF440)</li> <li>• Multi Frequency module (MF440)</li> <li>• Speech from hard-drive (SFH440)</li> <li>• SISI test</li> <li>• Master Hearing Aid (MHA440),</li> <li>• Hearing Loss Simulator (HLS440)</li> <li>• Loudness Scaling (LS440)</li> <li>• QuickSIN</li> <li>• TEN test</li> </ul>	<p><b>Included parts:</b></p> <ul style="list-style-type: none"> <li>• Affinity2.0 REM440 CD</li> <li>• OtoAccess™ database CD</li> <li>• IHM60 In-situ headset with probe microphone and reference microphone (double)</li> <li>• Probe tubes, 36 pcs.</li> <li>• Standard USB cable</li> <li>• Power cable 120 or 230V</li> <li>• Mouse pad</li> <li>• Instructions for Use document</li> </ul> <p><b>Optional parts:</b></p> <ul style="list-style-type: none"> <li>• SPL60 Transducer kit for RECD measurement including probes and eartips</li> <li>• BET60 Box with eartips for for RECD measurement.</li> <li>• Calibration adaptor for in-situ reference</li> <li>• VSP440 Visible Speech Mapping module</li> <li>• Optical USB 1.1 isolation extension cable</li> <li>• ACC60 Affinity2.0 carrying case</li> <li>• Coupler microphone extension cable</li> </ul>	<p><b>Included parts:</b></p> <ul style="list-style-type: none"> <li>• Affinity2.0 HIT440 CD</li> <li>• OtoAccess™ database CD</li> <li>• 2cc coupler with microphone and adaptors for ITE, BTE and Body Style HA</li> <li>• Coupler seal wax</li> <li>• Reference microphone</li> <li>• Standard USB cable</li> <li>• Power cable 120 or 230V</li> <li>• Mouse pad</li> <li>• Additional Information and Instructions for Use</li> </ul> <p><b>Optional parts:</b></p> <ul style="list-style-type: none"> <li>• Battery adapters BAA675, BAA13, BAA312, BAA10, BAA5</li> <li>• Couplers 1.2CC and 0.6CC: ITE, BTE, Ear simulator</li> <li>• TBS25M External test chamber incl. cables</li> <li>• ACC60 Affinity2.0 carrying case</li> <li>• Calibration adaptor</li> <li>• Optical USB 1.1 isolation extension cable</li> <li>• SKS10 Skull Simulator with power supply</li> </ul>

# General Technical Specifications

## Affinity<sup>2.0</sup> / Equinox<sup>2.0</sup> Hardware - Technical Specifications

<b>Medical CE-mark:</b>	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC Approval of the quality system is made by TÜV – identification no. 0123.	
<b>Safety Standards</b>	IEC 60601-1, UL60601-1, CAN/CSA-C22.2 No.60601-1 Class I, Applied parts type B, Continuous operation	
<b>EMC Standard</b>	IEC 60601-1-2 IEC 60645-1	
<b>Calibration</b>	Technical information is located in the specifications for the software modules. Calibration information and instructions are located in the Service manual.	
<b>PC requirements:</b>	2 GHz Intel Core 2 Duo CPU 2GB Ram 1.5 GB available disk space 1024x768 resolution (1280x1024 or higher recommended) Hardware accelerated DirectX/Direct3D graphics card. One or more USB ports, version 1.1 or higher. DVD-Rom drive.	
<b>Operative System:</b>	Windows 7, Windows 8, Windows 10	
<b>Display:</b>	1024x768 resolution (1280x1024 or higher recommended) Hardware accelerated DirectX/Direct3D graphics card.	
<b>Disc Space:</b>	1.5 GB available disk space	
<b>Compatible software</b>	Noah 3.7, Noah 4., OtoAccess™ and XML compatible Affinity <sup>2.0</sup> / Equinox <sup>2.0</sup> Suite	
<b>Input Specifications</b>	<b>Talk Back</b>	330 $\mu$ Vrms at max. input gain for 0dB VU-reading Input impedance : 47.5K $\Omega$
	<b>Mic. 1/TF &amp; Mic. 2</b>	
	<b>Pat. Resp. L &amp; R</b>	Switches 3.3V to the logic input. (The switch current is 33 $\mu$ A)
	<b>Inp. Aux. 1 &amp; 2</b>	20mVrms at max. input gain for 0dB VU-reading Input impedance : 15K $\Omega$
	<b>TB Coupler</b>	
	<b>TB Coupler - internal TB (Affinity2.0<sup>.0</sup> only)</b>	
	<b>Insitu L &amp; R - Probe mic.</b>	
	<b>CD1 &amp; CD2</b>	10mVrms at max input gain for 0dB VU-reading Input impedance : 10k $\Omega$
	<b>TB Ref.</b>	7mVrms at max. input gain for 0dB VU-reading Input impedance : 4,3K $\Omega$
	<b>TB Ref – internal TB (Affinity2.0<sup>.0</sup> only)</b>	
	<b>Insitu L &amp; R - Ref. mic</b>	
	<b>Ref.Mic./Ext.</b>	Not in use
	<b>Coupler/Ext.</b>	
	<b>Wave files</b>	Plays wave file from hard disk drive
<b>Output Specifications</b>	<b>FF1 / FF2 (Terminal Block)</b>	Up to 12.6Vrms by 8 $\Omega$ load 70Hz-20kHz $\pm$ 3dB
	<b>TB Lsp.</b>	
	<b>FF1/ FF2</b>	Up to 7Vrms by 600 $\Omega$ load 70Hz-20kHz $\pm$ 3dB
	<b>Sp 1, Sp 2, Sp 3, Sp 4</b>	
	<b>Left, Right</b>	Up to 7.0Vrms by 10 $\Omega$ load 70Hz-20kHz $\pm$ 3dB
	<b>Ins. Left, Ins. Right</b>	
	<b>Bone</b>	
	<b>Ins. Mask.</b>	
	<b>HF/HLS</b>	
	<b>Insitu L, Insitu R</b>	
	<b>Monitor, Ass. Mon.</b>	Max.3.5Vrms. by 8 $\Omega$ load

	<b>Sp. 1-4 Power Out</b>	70Hz-20kHz $\pm$ 3dB
	<b>DC</b>	Voltage: 5VDC Current: 0.5A
	<b>TB Loop</b>	Up to 100mA/meter
	<b>FF Loop</b>	70Hz-20kHz $\pm$ 3dB
	<b>Batt. Sim.</b>	Voltage: 1.1 – 1.6VDC Impedance range: 0 – 25 $\Omega$ .
	<b>Batt. Sim. - Internal TB (Affinity2.0<sup>.0</sup> only)</b>	
<b>Data Connections</b>	<b>USB/PC</b>	USB B socket for connection to PC (compatible with USB 1.1 and later)
	<b>USB</b>	USB A socket for connection of other USB devices (Internal USB 1.1 hub)
	<b>Keyb.</b>	Serial Peripheral Interface Bus (SPI interface) Check the Service manual for more information.
<b>Internal test box:</b>	Built in test box holds telecoil drive as well as special dual speaker set for checking directional microphone function.	
<b>Supported Systems</b>	Windows <sup>®</sup> 7 (32 and 64 bit) Windows <sup>®</sup> 8 (32 and 64 bit) Windows <sup>®</sup> 10 (32 and 64 bit)	
<b>Dimensions (LxWxH)</b>	Affinity2.0 <sup>.0</sup> : 42 x 38 x 14 cm / 16.5 x 15 x 5.5 inches	
<b>Weight</b>	Affinity2.0 <sup>.0</sup> : 5.5 kg / 12.1 lbs.	
<b>Power supply</b>	100-240 V~, 50-60Hz	
<b>Power Consumption:</b>	195VA	
<b>Operation environment</b>	Temperature:	15-35°C
	Re. Humidity:	30-90% Non condensing
<b>Transport and storage</b>	Transport temperature:	-20-50°C
	Storage temperature:	0-50°C
	Re. Humidity:	10-95% Non condensing

### Technical Specifications of the AC440 Software

<b>Medical CE-mark:</b>	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC. Approval of the quality system is made by TÜV – identification no. 0123.	
<b>Audiometer Standards:</b>	Tone: IEC60645-1/ANSI S3.6 Type 1 Speech: IEC60645-2/ANSI S3.6 Type A or A-E	
<b>Transducers &amp; Calibration:</b>	Calibration information and instructions are located in the Service manual. Check the accompanying Appendix for RETSPL levels for transducers	
<b>Air Conduction</b>		
DD45	PTB/DTU report 2009	Headband Static Force 4.5N ±0.5N
TDH39	ISO 389-1 1998, ANSI S3.6-2010	Headband Static Force 4.5N ±0.5N
HDA300	ISO 389-8 2006, ANSI S3.6-2010	Headband Static Force 8,8N ±0.5N
HDA280	PTB report 2004	Headband Static Force 5N ±0.5N
E.A.R Tone 3A/5A	ISO 389-2 1994, ANSI S3.6-2010	
IP30	ISO 389-2 1994, ANSI S3.6-2010 DES-2361	
CIR 33	ISO 389-2	
<b>Bone Conduction</b>	Placemenet: Mastoid	
B71	ISO 389-3 1994, ANSI S3.6-2010	Headband Static Force 5.4N ±0.5N
B81	ISO 389-3 1994, ANSI S3.6-2010	Headband Static Force 5.4N ±0.5N
<b>Free Field</b>	ISO 389-7 2005, ANSI S3.6-2010	
<b>High Frequency</b>	ISO 389-5 2004, ANSI S3.6-2010	
<b>Effective masking</b>	ISO 389-4 1994, ANSI S3.6-2010	
<b>Patient Response switch:</b>	Hand held push button.	
<b>Patient communication:</b>	Talk Forward and Talk Back.	
<b>Monitor:</b>	Output through external earphone or speaker.	
<b>Stimuli:</b>	Pure tone, Warble tone, NB, SN, WN, TEN noise	
<b>Tone</b>	125-20000Hz separated in two ranges 125-8000Hz and 8000-20000Hz. Resolution 1/2-1/24 octave.	
<b>Warble Tone</b>	1-10 Hz sine +/- 5% modulation	
<b>Wave file</b>	44100Hz sampling, 16 bits, 2 channels	
<b>Masking</b>	Automatic selection of narrow band noise (or white noise) for tone presentation and speech noise for speech presentation.	
Narrow band noise:	IEC 60645-1:2001, 5/12 Octave filter with the same centre frequency resolution as pure Tone.	
White noise:	80-20000Hz measured with constant bandwidth	
Speech Noise.	IEC 60645-2:1993 125-6000Hz falling 12dB/octave above 1KHz +/-5dB	
<b>Presentation</b>	Manual or Reverse. Single or multiple pulses. pulse time adjustable from 200mS-5000mS in 50mS steps. Simultaneous or alternating.	
<b>Intensity</b>	Check the accompanying Appendix for maximum output levels	
Steps	Available Intensity Steps is 1, 2 or 5dB	
Accuracy	Sound pressure levels: ± 2 dB. Vibration force levels: ± 5 dB.	
Extended range function	If not activated, the Air Conduction output will be limited to 20 dB below maximum output.	
<b>Frequency</b>	Range: 125Hz to 8kHz (Optional High Frequency: 8 kHz to 20 kHz) Accuracy: Better than ± 1 %	
<b>Distortion (THD)</b>	Sound pressure levels: below 1.5 % Vibration force levels: below 3 %.	
<b>Signal Indicator(VU)</b>	Time weighting: 350mS Dynamic range: -20dB to +3dB Rectifier characteristics: RMS Selectable inputs are provide with an attenuator by which the level can be adjusted to the indicator reference position(0dB)	
<b>Storing capability:</b>	Tone audiogram: dB HL, MCL, UCL, Tinnitus, R+L Speech Audiogram: WR1, WR2, WR3, MCL, UCL, Aided, Unaided, Binaural, R+L.	
<b>Compatible Software:</b>	Noah 4, Noah 3.7, OtoAccess™ and XML compatible	

**Technical Specifications - REM440 Software**

<b>Medical CE-mark:</b>	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC. Approval of the quality system is made by TÜV – identification no. 0123.	
<b>Real Ear Measurement Standards:</b>	IEC 61669, ISO 12124, ANSI S3.46.	
<b>Stimuli:</b>	Warble Tone, Pure Tone, Random noise, Pseudo random noise, Band limited white noise, Chirp, ICRA, Real Speech, any other sound file (automatic calibration available).	
<b>Frequency range:</b>	100Hz – 10kHz	
<b>Frequency accuracy:</b>	Less than $\pm 1\%$	
<b>Distortion:</b>	Less than 2%	
<b>Intensity range:</b>	40 – 90 dB	
<b>Intensity accuracy:</b>	Less than $\pm 1.5\%$	
<b>Measurement Intensity Range:</b>	Probe microphone 40-145 dB SPL $\pm 2$ dB.	
<b>Frequency Resolution:</b>	1/3, 1/6, 1/12, 1/24 octave or 1024 point FFT.	
<b>Probe microphone:</b>	Intensity: 40 – 140 dB	
<b>Reference microphone:</b>	Intensity: 40 – 100 dB	
<b>Intensity Accuracy:</b>	Less than $\pm 1.5$ dB	
<b>Cross talk</b>	Cross talk in the probe and probe tube will alter the obtained results with less than 1 dB at all frequencies.	
<b>Available tests:</b>	REUR REIG RECD REAR REAG	REORREOG REUG Input – Output FM Transparency Ear Level, FM only
<b>Compatible Software:</b>	Noah 4, Noah 3.7, OtoAccess™ and XML compatible	

## HIT440 Software - Technical Specifications

<b>Medical CE-mark:</b>	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC. Approval of the quality system is made by TÜV – identification no. 0123.	
<b>Hearing Aid Analyzer Standards:</b>	IEC 60118-0, IEC 60118-7, ANSI S3.22.	
<b>Frequency Range:</b>	100-10000Hz.	
<b>Frequency Resolution:</b>	1/3, 1/6, 1/12 and 1/24 octave or 1024 point FFT.	
<b>Frequency Accuracy:</b>	Less than $\pm 1\%$	
<b>Stimulus Signal:</b>	Warble Tone, Pure Tone, Random noise, Pseudo random noise, Band limited white noise, Chirp, ICRA, Real Speech, any other sound file (automatic calibration available).	
<b>Sweep Speed:</b>	1,5 – 12 sec.	
<b>FFT:</b>	Resolution 1024 points. Averaging: 10 – 500.	
<b>Stimulation Intensity Range:</b>	40-100 dB SPL in 1 dB step.	
<b>Intensity Accuracy:</b>	Less than $\pm 1.5$ dB	
<b>Measurement Intensity Range:</b>	Probe microphone 40-145 dB SPL $\pm 2$ dB.	
<b>Stimulus Distortion:</b>	Less than 1 % THD.	
<b>Battery Simulator:</b>	Standard and custom types are selectable	
	<i>Standard battery</i>	<i>Impedance[<math>\Omega</math>]</i> <i>Voltage[V]</i>
	Zinc air 5	8      1.3
	Zinc air 10	6      1.3
	Zinc air 13	6      1.3
	Zinc air 312	6      1.3
	Zinc air 675	3.5      1.3
	Mercury 13	8      1.3
	Mercury 312	8      1.3
	Mercury 657	5      1.3
	Mercury 401	1      1.3
	Silver 13	10      1.5
	Silver 312	10      1.5
	Silver 76	5      1.5
	Custom types	0 – 25      1.1 – 1.6
<b>Available tests:</b>	Additional tests can be designed by user	
	OSPL90 Full On Gain Input/Output Attack/Recovery Time Reference Test Gain Frequency Response Equivalent Input Noise	Harmonic Distortion Intermodulation Distortion Battery Current Drain Microphone Directionality Coil Frequency Response Coil Harmonic Distortion Coil Full-On Gain Response
<b>Pre-Programmed Protocols:</b>	HIT440 software comes with a set of Test Protocols loaded. Additional Test Protocols can be designed by user, or easily imported into the system.	
<b>Compatible Software:</b>	Noah 3.7, Noah 4., OtoAccess™ and XML compatible	